



Maxim Biomedical, Inc.

1500 E. Gude Drive, Suite A

Rockville, MD 20850

Tel: 301.251.0800

Fax: 301.217.9080

Web: www.mbi diagnostic.com

CAMBRIDGE BIOTECH HIV-1 SERUM WESTERN BLOT KIT

Part Number: 98002

Master Kit Lot: A0188

Certificate of Analysis

KIT EXPIRATION: 07/07/10

KIT COMPONENTS

	Lot No.	Exp. Date
Master Strip Lot	91061	07/16/10
Secondary Strip Lot	91041	07/14/10
Secondary Strip Lot	90971	07/07/10
Wash Buffer	WB369	07/29/10
Blotting Buffer	BB340	09/08/10
Blotting Powder	BP273	03/17/11
Conjugate 1	BG451	10/24/10
Conjugate 2	AH305	09/11/10
Substrate A	SA302	08/14/10
Substrate B	SB304	08/04/10
Non-Reactive Control	NC530	08/26/10
Weakly Reactive Control	WP333	09/19/10
Strongly Reactive Control	SP338	09/19/10

TESTING RESULTS

TEST

TEST	SPECIFICATION	
Physical Inspection	Kit must contain all components and evidence correct labeling and product insert	PASS
HIV-1 Western Blot Kit Response against Reference Panel & CBER	Must be equivalent to designations when in both the overnight and one-day assay format.	PASS
Infectivity of Lysed Antigen	Negative	PASS
Function Assay: consistency	Antigen profile must exhibit reactivity to p17, p24, p31, gp41, p51, p55, p66, and gp160.	PASS
Function Assay: specificity	Reference weak positive must exhibit reactivity to p24 and gp160	PASS

BIOBURDEN TESTING

COMPONENT

COMPONENT	SPECIFICATION	
Strips	No observable growth beyond 10 ⁻³ dilution	PASS
Wash Buffer	≤9 CFU / sample	PASS
Blotting Buffer	≤9 CFU / sample	PASS
Blotting Powder	≤9 CFU / gram	PASS
Conjugate 1	≤9 CFU / sample	PASS
Conjugate 2	≤9 CFU / sample	PASS
Substrate A	≤9 CFU / sample	PASS
Substrate B	≤9 CFU / sample	PASS
Non -Reactive Control	≤9 CFU / sample	PASS
Weakly Reactive Control	≤9 CFU / sample	PASS
Strongly Reactive Control	≤9 CFU / sample	PASS

Comments: All testing conducted and reported on this protocol passed all required specification.

Released By: Jan Nuff 08/17/09